

# CERTIFICATE OF CONFORMITY



**Applicant:** Shenzhen New Huayi Instrument Co., Ltd  
**Address:** F3, Block 2, Instrument World Industrial Park, Guiyue Road, Longhua New District, Shenzhen City  
**Product:** DIGITAL POWER CLAMP METER  
**Brand Name:** PEAKMETER  
**Model No.:** MS2203,MS2205  
**Ratings:** 4x1.5Vdc AA battery, 1000V CAT III, Max.1000A, class II  
**Test Report No.:** 16LTS030921E

The submitted sample of the above product has been tested for CE marking according to following European Directive and standards:

- Low Voltage Directive 2014/35/EU

**Standards:** EN61010-1:2010, EN61010-2-030:2010, EN61010-2-032:2012

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified European Directive. This verification does not imply assessment of the production of the product. The CE marking may be affixed if all relevant and effective European Directives with CE are applicable.



Vincent Fan  
Safety Manager  
March 16, 2016

This document shall not be reproduced, except in full, without the written approval of LiTest Technology Service Co., Ltd. Information given in this document is related to the tested specimen of the described electrical sample.

**LiTest Technology Service Co., Ltd**  
No.42 FuKang Road, HouJie Town,  
DongGuan City, GuangDong Province, China

Tel: +86-769-82272566/7  
Fax: +86-769-82272565  
Web: www.litest.cn Email:service@litest.cn