

CERTIFICATE OF CONFORMITY



Applicant: Shenzhen New Huayi Instrument Co., Ltd
Address: F3, Block 2, Instrument World Industrial Park, Guiyue Road, Longhua New District, Shenzhen City
Product: Motor and Phase Rotation Indicator
Brand Name: PEAKMETER
Model No.: MS5900
Ratings: 3x1.5Vdc AA battery, 600V CAT III, class II
Test Report No.: 16LTS030927E

The submitted sample of the above product has been tested for CE marking according to following European Directive and standards:

- Low Voltage Directive 2014/35/EU

Standards: EN61010-1:2010, EN61010-2-030:2010

The referred test report(s) show that the product complies with standard(s) recognized as giving presumption of compliance with the essential requirements in the specified European Directive. This verification does not imply assessment of the production of the product. The CE marking may be affixed if all relevant and effective European Directives with CE are applicable.



Vincent Fan
Safety Manager
March 18, 2016

This document shall not be reproduced, except in full, without the written approval of LiTest Technology Service Co., Ltd. Information given in this document is related to the tested specimen of the described electrical sample.

LiTest Technology Service Co., Ltd
No.42 FuKang Road, HouJie Town,
DongGuan City, GuangDong Province, China

Tel: +86-769-82272566/7
Fax: +86-769-82272565
Web: www.litest.cn Email:service@litest.cn